

Withdrawal of Weian Type IIR Surgical Masks

We have been made aware that a Type IIR facemask produced by Weian may contain latex and could cause anaphylactic shock in anyone who is allergic to latex. The recommendation is to avoid wearing these masks and use an alternative brand.

Please see the photographs below to help with the identification of the masks.



Withdrawal of Gloves – Do Not Use these Specifications

The following gloves are not fit for use because they do not meet EN455 rating.



The specification for the following items cannot be validated:

